

ASB Standard 017, Second Edition  
2024

**Standard for Metrological Traceability in Forensic  
Toxicology**

DRAFT



**ASB**  
**ACADEMY**  
**STANDARDS BOARD**

## Standard for Metrological Traceability in Forensic Toxicology

ASB Approved Xxxxx 2024

ANSI Approved Xxxxx 2024



**ASB**  
ACADEMY  
STANDARDS BOARD

410 North 21<sup>st</sup> Street  
Colorado Springs, CO 80904

This document may be downloaded from: [www.aafs.org/academy-standards-board](http://www.aafs.org/academy-standards-board)

*This document is provided by the AAFS Academy Standards Board. Users are permitted to print and download the document and extracts from the document for personal use, however the following actions are prohibited under copyright:*

- *modifying this document or its related graphics in any way;*
- *using any illustrations or any graphics separately from any accompanying text; and,*
- *failing to include an acknowledgment alongside the copied material noting the AAFS Academy Standards Board as the copyright holder and publisher.*

*Users may not reproduce, duplicate, copy, sell, resell, or exploit for any commercial purposes this document or any portion of it. Users may create a hyperlink to [www.aafs.org/academy-standards-board](http://www.aafs.org/academy-standards-board) to allow persons to download their individual free copy of this document. The hyperlink must not portray AAFS, the AAFS Standards Board, this document, our agents, associates and affiliates in an offensive manner, or be misleading or false. ASB trademarks may not be used as part of a link without written permission from ASB.*

*The AAFS Standards Board retains the sole right to submit this document to any other forum for any purpose.*

*Certain commercial entities, equipment or materials may be identified in this document to describe a procedure or concept adequately. Such identification is not intended to imply recommendations or endorsement by the AAFS or the AAFS Standards Board, nor is it intended to imply that the entities, materials, or equipment are necessarily the best available for the purpose.*

*Proper citation of ASB documents includes the designation, title, edition, and year of publication.*

*This document is copyrighted © by the AAFS Standards Board, LLC. 2024. All rights are reserved.  
410 North 21st Street, Colorado Springs, CO 80904, [www.aafs.org/academy-standards-board](http://www.aafs.org/academy-standards-board)*

## Foreword

This Standard was developed to provide minimum requirements for establishing metrological traceability in forensic toxicology. Establishing traceability of a measurement ensures confidence and reliability in forensic toxicological test and calibration results.

This 2<sup>nd</sup> Edition includes several substantive changes from the 1<sup>st</sup> Edition. “Measurement Traceability” was replaced with “Metrological Traceability” in the title of this standard and throughout the document. The scope was revised to clarify the subdisciplines of forensic toxicology for which the standard is applicable. ANSI/ASB Standard 054, *Standard for a Quality Control Program in Forensic Toxicology Laboratories* and ANSI/ASB Standard 056, *Standard for Evaluation of Measurement Uncertainty in Forensic Toxicology* were included as additional normative references. Some terms and definitions were modified to apply to both forensic toxicology testing and calibration of alcohol measuring instruments. The Calibration Program section from the 1<sup>st</sup> Edition was determined to not be within the scope of the document, so this section was removed in this edition. Changes were made to the section title of Certified Reference Materials, and a requirement was added to address situations when a CRM is unavailable. The Reference Standards section title was also changed, and the mass reference standard calibration frequency requirement was revised. The requirement for calibration of rulers was moved to its own section and revised. The requirement for calibration of thermometers used to verify equipment performance and not to establish metrological traceability was removed from the document. Finally, the title of the Conformance section was changed, and the requirement was revised.

The American Academy of Forensic Sciences established the Academy Standards Board (ASB) in 2015 with a vision of safeguarding Justice, Integrity and Fairness through Consensus Based American National Standards. To that end, the ASB develops consensus-based forensic standards within a framework accredited by the American National Standards Institute (ANSI), and provides training to support those standards. ASB values integrity, scientific rigor, openness, due process, collaboration, excellence, diversity, and inclusion. ASB is dedicated to developing and making freely accessible the highest quality documentary forensic science consensus Standards, Guidelines, Best Practices, and Technical Reports in a wide range of forensic science disciplines as a service to forensic practitioners and the legal system.

The Toxicology Subcommittee of the Organizational Scientific Area Committee developed the first version of this document. That version and this subsequent revision were prepared and finalized as a standard by the Toxicology Consensus Body of the ASB.

Questions, comments, and suggestions for the improvement of this document can be sent to AAFS-ASB Secretariat, [asb@aafs.org](mailto:asb@aafs.org) or 401 N 21<sup>st</sup> Street, Colorado Springs, CO 80904.

All hyperlinks and web addresses shown in this document are current as of the publication date of this standard.

ASB procedures are publicly available, free of cost, at [www.aafs.org/academy-standards-board](http://www.aafs.org/academy-standards-board).

**Keywords:** *Metrological Traceability, Calibration, Forensic Toxicology*

**Table of Contents** *(to be finalized upon completion)*

1	Scope .....	1
2	Normative References .....	1
3	Terms and Definitions .....	1
4	Background .....	3
5	Requirements for Metrological Traceability .....	4
5.1	General .....	4
5.2	Implementation of Metrological Traceability .....	5
5.3	Calibration Service Provider .....	5
5.4	Reference Materials .....	6
6	Document Retention.....	8
	Annex A (informative).....	9

DRAFT

# Standard for Metrological Traceability in Forensic Toxicology

## 1 Scope

This standard defines the minimum requirements for establishing metrological traceability in forensic toxicology. Specifically, it is intended for the subdisciplines of postmortem forensic toxicology, human performance toxicology (e.g., drug-facilitated crimes and driving-under-the-influence of alcohol or drugs), non-regulated employment drug testing, court-ordered toxicology (e.g., probation and parole, drug courts, child services), general forensic toxicology (non-lethal poisonings or intoxications) and calibration of breath alcohol measuring instruments.

## 2 Normative References

The following references are documents that are indispensable for the application of the standard. For dated references, only the edition cited applies. The latest edition of the referenced document (including any amendments) applies for undated references.

International Bureau of Weights and Measures (BIPM)-International Committee for Weights and Measures (CIPM) *Mutual Recognition Arrangement*<sup>1</sup>

International Laboratory Accreditation Cooperation (ILAC), ILAC P10:07/2020 *ILAC Policy on Metrological Traceability of Measurement Results*<sup>2</sup>

International Laboratory Accreditation Cooperation (ILAC) *Mutual Recognition Arrangement*<sup>2</sup>

International Organization for Standardization (ISO), ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories* (Geneva, Switzerland: ISO, 2017)

ANSI/ASB Standard 054, *Standard for a Quality Control Program in Forensic Toxicology Laboratories*.<sup>3</sup>

ANSI/ASB Standard 056, *Standard for Evaluation of Measurement Uncertainty in Forensic Toxicology*.<sup>3</sup>

## 3 Terms and Definitions

For purposes of this document, the following definitions and acronyms apply.

### 3.1 accreditation<sup>4</sup>

Third-party attestation of a forensic science service provider conveying formal demonstration of its competence, impartiality, and consistent operation in performing specific conformity assessment activities.

---

<sup>1</sup> More information about the BIPM is available at: <http://www.bipm.org/en/cipm-mra/>

<sup>2</sup> ILAC document available for download at: <https://ilac.org/?ddownload=123220>

<sup>3</sup> Available from: <https://www.aafs.org/academy-standards-board>

<sup>4</sup> International Organization for Standardization (ISO), ISO/IEC 17000:2020 Conformity assessment — Vocabulary and general principles (Geneva, Switzerland)

- 31 **3.2**  
32 **accuracy**<sup>5</sup> (Modified)  
33 Closeness of agreement between a measured quantity value and a true quantity value of a  
34 measurement.
- 35 **3.3**  
36 **calibration**<sup>5</sup> (Modified)  
37 Operation that, under specified conditions, establishes a relation between the quantity value and  
38 corresponding indications.
- 39 **3.4**  
40 **calibrator**<sup>5</sup>(Modified)  
41 Measurement standard used in calibration.
- 42 **3.5**  
43 **certified reference material**<sup>6</sup>  
44 **CRM**  
45 Reference material characterized by a metrologically valid procedure for one or more specified  
46 properties, accompanied by a certificate that provides the value of the specified property, its  
47 associated uncertainty, and a statement of metrological traceability.
- 48 **3.6**  
49 **control**  
50 Material of known composition used to evaluate the performance of a method.
- 51 **3.7**  
52 **decision point**  
53 Administratively defined cutoff or concentration that is at or above the method's limit of detection  
54 or lower limit of quantitation and is used to discriminate between a negative and positive test  
55 result.
- 56 **3.8**  
57 **forensic science service provider**  
58 Forensic science agency or forensic science practitioner providing forensic science services.
- 59 **3.9**  
60 **limit of detection**  
61 **LOD**  
62 Estimate of the lowest concentration of an analyte in a sample that can be reliably differentiated  
63 from blank matrix and meets identification criteria.

---

<sup>5</sup> Committee for Guides in Metrology (JCGM), International vocabulary of metrology – Basic and general concepts and associated terms (VIM), 3rd ed. (Sèvres, France)

<sup>6</sup> International Organization for Standardization (ISO), ISO Guide 30:2015 Reference Materials – Selected Terms and Definitions (Geneva, Switzerland)

**64 3.10****65 lower limit of quantitation****66 LLOQ**

67 Estimate of the lowest concentration of an analyte in a sample that can be reliably measured with  
68 acceptable bias and precision.

**69 3.11****70 measurement<sup>5</sup>**

71 Process of experimentally obtaining one or more quantity values that can reasonably be attributed  
72 to a quantity.

**73 3.12****74 metrological traceability<sup>5</sup>**

75 Property of a measurement result whereby the result can be related to a reference through a  
76 documented unbroken chain of calibrations, each contributing to the measurement uncertainty.

**77 3.13****78 quantity<sup>5</sup>**

79 Property of a phenomenon, body, or substance, where the property has a magnitude that can be  
80 expressed as a number and a reference.

**81 3.14****82 quantity value<sup>5</sup>**

83 Number and reference together expressing magnitude of a quantity.

**84 3.15****85 reference material<sup>6</sup>****86 RM**

87 Material, sufficiently homogeneous and stable with reference to specified properties, which has  
88 been established to be fit for its intended use in a measurement or in examination of nominal  
89 properties.

**90 3.16****91 reference standard<sup>5(Modified)</sup>**

92 Measurement standard that is used to calibrate or verify (working reference standard) measuring  
93 instruments or measuring systems.

**94 4 Background**

95 **4.1** Measurement relates to the entire process of obtaining a quantity value. Some measurement  
96 processes require only a single step, whereas others may require more than one step and are made  
97 in different ways in testing and calibration processes. A measurement may not be the reported  
98 result but may affect the validity of the reported result. Measurements may be used as the basis for  
99 an interpretation or opinion. Examples of measurements include:

- 100 a) a reported test result;
- 101 b) a reported calibration result;
- 102 c) a quantitative decision point (cutoff) that results in a qualitative test report; or



103 d) aliquoting a sample using a different pipette than that used for all calibrators and controls.

104 **4.2** Metrological traceability can be characterized by the following essential elements.<sup>7</sup>

105 a) *Unbroken Chain of Comparisons* – A documented system of comparisons with each step having  
106 the essential elements of metrological traceability going back to a stated reference acceptable to  
107 the parties, usually a national or international standard.

108 b) *Documented Measurement Uncertainty* – The measurement uncertainty for each step in the  
109 traceability chain must be calculated according to defined methods and must be stated so that  
110 an overall uncertainty for the whole chain may be calculated.

111 c) *Documented Measurement Procedure* – Each step in the chain must be performed according to  
112 documented, generally-accepted procedures and the results must be documented.

113 d) *Technical Competence* – The laboratories or bodies performing one or more steps in the chain  
114 must maintain and supply evidence of technical competence (e.g., by maintaining appropriate  
115 training records, participating in inter-laboratory comparisons, and by demonstrating that they  
116 are accredited by a recognized accreditation body).

117 e) *Realization of SI Units* – The chain of comparisons must, where possible, end at the realization of  
118 the International System of Units (SI).

119 f) *Documented Calibration Intervals* – Calibrations must be repeated at established and  
120 appropriate intervals to preserve metrological traceability.

121 g) *Measurement Assurance* – A proper measurement assurance program [however named] must  
122 be established to ensure the validity of the measurement process and to ensure the calibration  
123 status of equipment, reference standards and reference materials.

## 124 **5 Requirements for Metrological Traceability**

### 125 **5.1 General**

126 **5.1.1** Forensic science service providers shall establish traceability of a measurement process by:

127 a) making one or more measurements using equipment that has been calibrated with established  
128 metrological traceability and/or,

129 b) through the use of calibrators in the test or calibration method.

130 *NOTE 1* Calibrators may be undiluted CRM(s) or calibrators prepared using calibrated equipment and CRMs  
131 or non-certified RMs shown to be fit for purpose.

132 *NOTE 2* Controls are not a way to establish metrological traceability. Metrologically-prepared controls  
133 provide information on both precision and bias.

---

<sup>7</sup> See “GMP 13 Good Measurement Practice for Ensuring Metrological Traceability”  
[GMP 13 Ensuring Metrological Traceability \(nist.gov\)](https://www.nist.gov/gmp13)



134 **5.1.2** Proper handling and storage procedures that meet or exceed the manufacturer's  
135 recommendations shall be followed for equipment, certified reference materials, and non-certified  
136 reference materials used to establish and maintain metrological traceability.

## 137 **5.2 Implementation of Metrological Traceability**

138 **5.2.1** The measurements in a method that require metrological traceability shall be identified  
139 during the method development process.

140 *NOTE This may be only the calibrators used, or it may be calibrators and one or more additional measurements*  
141 *made during the testing or calibration method.*

142 **5.2.2** The manner in which metrological traceability will be established shall be determined  
143 during the method development process.

144 **5.2.3** Metrological traceability shall be established during method validation and routine use of:

145 a) qualitative methods with a decision point (quantitative) calibrator

146 *Examples:*

147 1) *immunoassay method with a cutoff concentration*

148 2) *chromatography-based method with a decision point concentration*

149 b) quantitative methods utilizing a calibration curve, and

150 c) breath alcohol instrument calibration methods.

## 151 **5.3 Calibration Service Provider**

152 **5.3.1** Calibration of the equipment used to make the measurement, including equipment  
153 purchased with a calibration certificate, shall be performed by an accredited calibration service  
154 supplier, if available, that is either:

155 a) a National Metrology Institute (NMI) that is a signatory to the BIPM - CIPM *Mutual Recognition*  
156 *Arrangement*, with the calibration to be performed listed in Appendix C of the BIPM *Key*  
157 *Comparison Database* (KCDB); or

158 b) a service supplier accredited to ISO/IEC 17025:2017 by an accrediting body that is a signatory  
159 to the International Laboratory Accreditation Cooperation (ILAC) *Mutual Recognition*  
160 *Arrangement*, with the calibration to be performed listed in a scope of accreditation.

161 **5.3.2** If equipment calibration is not available from a supplier that meets the above requirement,  
162 then the forensic science service provider shall:

- 163 a) perform an evaluation to ensure the supplier meets the requirements for competence,  
 164 metrological traceability, and measurement capability in ISO/IEC 17025:2017 or in the *ILAC Policy*  
 165 *on the Traceability of Measurement Results*<sup>8</sup>;
- 166 b) keep objective evidence of this evaluation; and
- 167 c) perform a re-evaluation of services at least every two years.

## 168 5.4 Reference Materials

169 **5.4.1** A CRM, if available, shall be obtained by the forensic science service provider from a  
 170 supplier that is either:

- 171 a) a National Metrology Institute (NMI) that is a signatory to the BIPM – CIPM *Mutual Recognition*  
 172 *Arrangement*, with the CRM to be purchased included in the BIPM key comparison database  
 173 (KCDB); or
- 174 b) an accredited Reference Material Producer that is accredited to ISO 17034:2016 by an  
 175 accrediting body that is a signatory to a mutual or multilateral recognition arrangement in an  
 176 ILAC-recognized regional accreditation cooperation or the *ILAC Mutual Recognition*  
 177 *Arrangement*, with a scope of accreditation covering the CRM.

178 **5.4.2** If a CRM is not available from a supplier that meets the above requirement, then the  
 179 forensic science service provider shall:

- 180 a) perform an evaluation to:
- 181 1) ensure the supplier is accredited to ISO/IEC 17025:2017 by an accrediting body that is a  
 182 signatory to a mutual or multilateral recognition arrangement in an ILAC-recognized  
 183 regional accreditation cooperation or the *ILAC Mutual Recognition Arrangement*, with a  
 184 scope of accreditation covering the CRM; or,
- 185 2) ensure the supplier and the CRM meet the requirements for competence, metrological  
 186 traceability, and measurement capability in ISO/IEC 17025:2017 or in the *ILAC Policy on*  
 187 *the Traceability of Measurement Results*.<sup>11</sup>
- 188 b) keep objective evidence of this evaluation; and
- 189 c) perform a re-evaluation of the supplier at least every two years.

190 **5.4.3** If a CRM is not available and a non-certified RM is used to prepare a calibrator, then:

- 191 a) the RM shall be evaluated as required in ANSI/ASB Standard 054, *Standard for a Quality Control*  
 192 *Program in Forensic Toxicology Laboratories*; and
- 193 b) the measurement uncertainty of the RM quantitative value shall be determined as required in  
 194 ANSI/ASB Standard 056, *Standard for Evaluation of Measurement Uncertainty in Forensic*  
 195 *Toxicology*.

<sup>8</sup> Reference to ILAC P10 <https://ilac.org/?ddownload=123220/>

196 **5.4.4** Whether certified or not, if a RM is diluted (i.e., preparing a stock or working solution) then  
197 the equipment used shall be calibrated per section 5.3.

## 198 **5.5 Equipment**

### 199 **5.5.1 General**

200 **5.5.1.1** A forensic toxicology testing or calibration laboratory shall have a procedure that:

- 201 a) identifies which equipment requires calibration;
- 202 b) requires equipment to be calibrated prior to use;
- 203 c) includes the frequency of calibration and acceptability/tolerance specifications for equipment;
- 204 d) requires equipment to be calibrated to include the range of use;
- 205 e) includes an evaluation of the need for intermediate checks based on risk considering, but not  
206 limited to, the frequency of use, work volume, occurrence of unexpected shutdown, and  
207 equipment maintenance; and
- 208 f) specifies the actions to be taken when a calibration or intermediate check does not meet  
209 acceptability/tolerance specifications.

210 **5.5.1.2** If intermediate checks are performed, the procedure shall:

- 211 a) require use of calibrated equipment (e.g., mass reference standards, equipment used to monitor  
212 environmental conditions); and
- 213 b) include the frequency and acceptability/tolerance specifications for intermediate checks.

### 214 **5.5.2 Calibration of Analytical Equipment**

#### 215 **5.5.2.1 General**

216 The equipment in 5.5.2.2 through 5.5.2.7 requires calibration when used to establish metrological  
217 traceability and/or when the measurement accuracy affects the validity of the reported result.

#### 218 **5.5.2.2 Analytical Balances**

219 Analytical balances shall be calibrated at least annually.

#### 220 **5.5.2.3 Mass Reference Standards**

221 Mass reference standards used to verify the accuracy of equipment shall be:

- 222 a) calibrated at least once every two years;<sup>9</sup>

---

<sup>9</sup> GMP 11 “Good Measurement Practice for Assignment and Adjustment of Calibration Intervals for Laboratory Standards” [NIST GMP 11](#)

223 b) adjusted by an accredited calibration service supplier, and the reference standards shall be  
224 calibrated before and after any adjustment; and

225 c) dedicated for this purpose unless the forensic toxicology laboratory has demonstrated that  
226 their integrity as reference standards are maintained.

#### 227 **5.5.2.4 Volumetric Glassware**

228 Class A volumetric glassware used to prepare calibrators shall be calibrated at least once every ten  
229 years.<sup>10</sup>

#### 230 **5.5.2.5 Pipettes, Diluters, and Syringes**

231 All pipettes, pipette diluters, automatic diluters, and syringes used to prepare calibrator solutions  
232 that require metrological traceability or in sample preparation (e.g., sample aliquoting and other  
233 steps that affect overall measurement uncertainty) shall be calibrated at least annually.

234 *NOTE Autosampler syringes used for sample introduction to analytical instrumentation (e.g., gas*  
235 *chromatograph, liquid chromatograph, or immunoassay) do not require calibration.*

#### 236 **5.5.2.6 Rulers**

237 Rulers shall be calibrated at least once every ten years.

#### 238 **5.5.2.7 Breath Alcohol Calibration Equipment**

239 **5.5.2.7.1** Simulator thermometers, multimeters, and barometers, as applicable, shall be  
240 calibrated at least every two years.

241 **5.5.2.7.2** Sections 5.5.2.2 through 5.5.2.5 also apply to equipment used in the preparation of  
242 breath alcohol measurement standards used in a breath alcohol calibration method.

#### 243 **5.5.3 Other Equipment**

244 General laboratory equipment used during sample preparation (e.g., centrifuges, rotators, shakers,  
245 water baths, evaporators, extraction manifolds, heating blocks, and pH meters) that does not  
246 significantly affect the validity of the reported result does not require calibration to establish  
247 metrological traceability. A forensic toxicology laboratory may elect to use calibration or  
248 verification as a maintenance procedure to ensure the proper functioning of the equipment.

### 249 **6 Document Retention**

250 Documentation to verify conformance with the above requirements shall be maintained for a length  
251 of time defined by the forensic science service provider.

252

---

<sup>10</sup> The National Institute of Standards and Technology (NIST) Important Technical Guidance on Glassware  
<https://www.nist.gov/system/files/documents/2017/05/09/h-008.pdf>

253  
254  
255

## Annex A (informative)

### Bibliography

256 The following bibliography is not intended to be an all-inclusive list, review, or endorsement of  
257 literature on this topic. The goal of the bibliography is to provide examples of publications  
258 addressed in the standard.

- 259 1] ASTM International E542-22 *Standard Practice for Gravimetric Calibration of Laboratory*  
260 *Volumetric Instruments*<sup>11</sup>
- 261 2] *Eurachem Terminology in Analytical Measurement – Introduction to VIM 3*<sup>12</sup>
- 262 3] The National Institute of Standards and Technology (NIST) definition of “internal  
263 measurement assurance program”<sup>13</sup>
- 264 4] International Organization for Standardization (ISO), ISO 17034:2016 *General requirements for*  
265 *the competence of reference material producers* (Geneva, Switzerland: ISO, 2016)<sup>14</sup>
- 266 5] International Organization for Standardization (ISO), ISO Guide 33: 2015 *Reference Materials –*  
267 *Good Practice in Using Reference Materials* (Geneva, Switzerland) ISO documents available for  
268 purchase from the ISO Standards Catalogue or from other authorized distributors<sup>14</sup>
- 269 6] International Bureau of Weights and Measures (BIPM) *Key Comparison Database (KCDB),*  
270 *Appendix C*<sup>15</sup>

---

<sup>11</sup> Available from: [www.astm.org](http://www.astm.org)

<sup>12</sup> Available from: <https://www.eurachem.org/index.php/publications/guides/terminology-in-analytical-measurement>

<sup>13</sup> Available from: <https://www.nist.gov/metrology/metrological-traceability>

<sup>14</sup> Available from: <http://www.iso.org/standards-catalogue/browse-by-ics.html>

<sup>15</sup> Available from: <https://www.bipm.org/en/cipm-mra/kcdb>

DRAFT



**ASB**  
ACADEMY  
STANDARDS BOARD

Academy Standards Board  
410 North 21st Street  
Colorado Springs, CO 80904  
[www.aafs.org/academy-standards-board](http://www.aafs.org/academy-standards-board)